Certificate of Analysis

YSI Incorporated
Yellow Springs Instrument Co., Yellow Springs, Ohio 45387 USA
Phone 937 767-7241•800 765-4974•Fax 937 767-9353

Product Category	Buffer and solid products
Product Model Number	2363
Lot Number	26G100021
Method	Verification of mass

Determination	Specification	Result
Average Mass	1.15-1.35 grams	1.25 g

The procedures used to manufacture and test this product have been verified by QA for compliance. The electronic signature below is representative of the Chemistry Technician testing and verifying conformity to acceptance criteria.

Completed by: Amanda Fenner Date: 7/10/2026

ISO 9001 CERTIFIED YSI INCORPORATED